

The detectors in the microscope typically last at least 10 hours. However, it is good practice to top off the detectors after 8 hours. Refilling the detectors while the synchrotron beam is being refilled is an easy way to remember!

[See the FAQ on refilling the 4L nitrogen dewar.](#)